



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development -
Environmental Team

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
03/03/2025	R250453	0.22	0.09	3.58	3.48	2.83	10	94.0	Pass
05/03/2025	R250464	0.22	0.08	3.29	3.42	-3.87	10	93.2	Pass
07/03/2025	R250502	0.22	0.08	3.89	3.79	2.60	10	93.1	Pass
10/03/2025	R250522	0.22	0.11	3.43	3.53	-2.87	10	93.9	Pass
12/03/2025	R250535	0.22	0.09	5.18	4.93	4.95	10	94.1	Pass
14/03/2025	R250549	0.22	0.11	3.50	3.64	-3.92	10	94.5	Pass
17/03/2025	R250591	0.22	0.07	4.30	4.16	3.31	10	93.6	Pass
19/03/2025	R250609	0.22	0.09	4.84	5.01	-3.45	10	94.5	Pass
21/03/2025	R250615	0.22	0.07	4.72	4.54	3.89	10	94.1	Pass
24/03/2025	R250627	0.22	0.08	4.84	4.98	-2.85	10	93.8	Pass
26/03/2025	R250657	0.22	0.11	3.81	3.68	3.47	10	93.8	Pass
28/03/2025	R250682	0.22	0.08	3.50	3.58	-2.26	10	93.3	Pass
31/03/2025	R250694	0.22	0.07	3.26	3.19	2.17	10	94.1	Pass



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Service Contract No. WD/03/2023 Hung Shui Kiu/Ha Tsuen New Development Area Second Phase Development -
Environmental Team

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
03/03/2025	R250454	0.22	0.09	3.46	3.59	-3.69	10	94.4	Pass
05/03/2025	R250465	0.22	0.11	5.07	4.88	3.82	10	92.9	Pass
07/03/2025	R250503	0.22	0.09	4.99	5.12	-2.57	10	92.7	Pass
10/03/2025	R250523	0.22	0.07	3.91	3.83	2.07	10	95.0	Pass
12/03/2025	R250536	0.22	0.10	4.06	4.19	-3.15	10	93.7	Pass
14/03/2025	R250550	0.22	0.09	5.30	5.48	-3.34	10	93.9	Pass
17/03/2025	R250592	0.22	0.11	3.87	3.71	4.22	10	93.0	Pass
19/03/2025	R250610	0.22	0.08	4.73	4.88	-3.12	10	94.3	Pass
21/03/2025	R250616	0.22	0.10	3.65	3.51	3.82	10	93.5	Pass
24/03/2025	R250628	0.22	0.09	4.84	5.02	-3.65	10	92.7	Pass
26/03/2025	R250658	0.22	0.07	3.34	3.22	3.66	10	94.3	Pass
28/03/2025	R250683	0.22	0.10	3.98	4.11	-3.21	10	93.7	Pass
31/03/2025	R250695	0.22	0.07	4.79	4.63	3.40	10	94.8	Pass